

FLEXILUX 200 LED

LED Light Source



- Lowest cost of ownership with new LED technology
 - No bulbs to exchange – long lifetime (> 20,000 h)
 - Low total power consumption (70 % less than a 180 W xenon light source)
- High luminance with low temperature development
- User friendly: integrated SCHÖLLY MIS-Bus interface offers direct connectivity to camera – brightness can be adjusted from camera head
- Outstanding performance
 - Light intensity dimmable 5 - 100 %
 - Large LCD display shows dim level – easy to read from a distance
 - Standby modus of the LED (5 %)
- Multi light guide adapter for Storz, Olympus and Wolf
- Low fan noise because of low temperature development

PRIVATE LABEL PRODUCTS - OUR EXPERTISE LABELED WITH YOUR NAME




Logo / Labeling



Front frame color



Button color



SCHÖLLY FIBEROPTIC GMBH
Robert-Bosch-Strasse 1-3
79211 Denzlingen, Germany

Phone: +49 7666 908-0
Fax: +49 7666 908-380

www.schoelly.de/ME
info@schoelly.de

FLEXILUX 200 LED

LED Light Source




TECHNICAL DATA

DESCRIPTION	ITEM NUMBER
LED light source Light guide, 4.8 mm fiber-Ø, 2.3 m length	05.0740led 05.0090l

TECHNICAL DATA	
Lamp	Type: LED Color temperature: approx. 6,500 Kelvin Lifetime: > 20,000 h Luminous flux: 1,400 lm
Operation control panel	On-Off with LED lamp LCD display with current dimming Button for illumination up / down Multi light guide adapter
Connectivity	SCHÖLLY MIS-Bus interface
Dimensions / Weight	295 x 130 x 355 mm (W x H x D) / 5.8 kg
Cleaning	Wiping with disinfectant
SYSTEM	
Standard compliance	IEC/DIN EN 60601-1 2 nd and 3 rd EN 62471
Electromagnetic compatibility	EN 60601-1-2
Directive compliance	CE-mark
According to medical directive 93/42 EEC class 1	
Power connection	100 - 240 V AC, 50/60 Hz, 1.0 - 0.42 A
Power consumption	Approx. 140 W

Native Full HD 1080p endoscopy camera for direct connection via MIS-Bus and other BF/CF light guides are available. Please contact us for further information.

DETAILS



Light guide connection



Light guide



MIS-Bus

Technical specifications are subject to change. The information is not legally binding. The contents are only for information about our products. Reprint, in whole or in part, is not permitted. Medical devices are only allowed to be marketed and used in countries where they have been approved for sales by regulatory authorities. For information on the available market access please contact us.

MEDICAL ENDOSCOPY
MPA302-000-01-ME
Version: 0
2018-Jan-22